



INSTRUCTIONS

N-Histofine[®]

MOUSESTAIN KIT

Immunohistochemical staining kit

Store at 2-8°C

Code : 414321F50 Slides

Code : 414322F500 Slides

Made in Japan

NICHIREI BIOSCIENCES INC.

6-19-20, Tsukiji, Chuo-ku, Tokyo, 104-8402, JAPAN

Phone:81-3-3248-2207Facsimile:81-3-3248-2243

1. INTRODUCTION
N-Histofine[®] MOUSESTAIN KIT is the mouse on mouse system. It is designed specifically to allow immunohistochemical staining with a mouse primary antibody on formalin-fixed paraffin-embedded mouse tissue sections. **N-Histofine[®]** MOUSESTAIN KIT uses **Universal Immuno-enzyme Polymer (UIP)** method which is the patented technology.

2. PRESENTATION
N-Histofine[®] MOUSESTAIN KIT consists of three kinds of reagents. Liquid. Ready to use.

components		Code	
		414321F	414322F
Vial 1	Blocking reagent A	50 Slides	500 Slides
Vial 2	Blocking reagent B	6mL	51mL (3 × 17mL)
Vial 3	Simple Stain Mouse MAX PO(M)	6mL	51mL (3 × 17mL)

Simple Stain Mouse MAX PO(M) is the labeled polymer prepared by combining amino acid polymers with peroxidase and goat anti-mouse Ig which is reduced to Fab . It is stored in MOPS (3-Morpholinopropanesulfonic acid) buffer (pH 6.5) containing stabilizer and antibacterial agent.

3. INTENDED USE
FOR RESEARCH USE ONLY.
N-Histofine[®] MOUSESTAIN KIT is designed specifically to allow immunohistochemical staining with a mouse primary antibody on formalin-fixed paraffin-embedded mouse tissue sections.

4. PRINCIPLE
The antigen / antibody / Universal Immuno-peroxidase Polymer complex can be prepared by allowing the reagent to react with a mouse primary antibody bound to the antigen on mouse tissue sections. The enzymatic activity of this complex results

in a colored deposit, thus staining the antigen site. To eliminate background staining, this kit uses Blocking reagent A and Blocking reagent B.

- 5. PRECAUTION WHILE USING OR HANDLING.**
- Before using this reagent, please read these instructions.
 - Do not use reagents after the expiration date.
 - Specimens, before and after fixation, and all other materials exposed to them, should be handled like biohazardous materials with proper precautions.
 - Inhalation or ingestion of the highly allergic fixative formaldehyde is harmful. Wear protective mask. If swallowed, induce vomiting. If skin or eye contact occurs, wash thoroughly with water.
 - Organic reagents are flammable. Do not use near open flame.
 - Never pipette reagents by mouth and avoid their contact with skin, mucous membranes and clothes.
 - Avoid microbial contamination of reagents as incorrect result may occur.
 - Avoid splashing of reagents or generation of aerosols.
 - For research use only. Not for diagnostic use.

- 6. STAINING PROCEDURES**
- Reagents and Materials required but not provided
- Mouse primary antibody
 - Xylene
 - 100% ethanol
 - 95% ethanol
 - Counter staining solution
 - 3% solution of hydrogen peroxide in absolute methanol (Add 1 part of 30% hydrogen peroxide to 9 parts of absolute methanol)
 - Phosphate buffered saline (PBS) (pH 7.6±0.2)
 - NaCl 7.75 g
 - K₂HPO₄ 1.50 g
 - KH₂PO₄ 0.20 g
 - distilled water 1L
 - Adhesive for tissue section (0.02% poly-L-lysine, silane or the like)
 - Chromogen/substrate reagent
 - **N-Histofine[®]** Simple Stain AEC solution (No preparation, One-bottle) 500 tests Code: 415182F 1500 tests Code: 415184F
- NICHIREI BIOSCIENCES developed One-bottle, ready to use and no preparation chromogen/substrate called **N-Histofine[®]** Simple Stain Substrate Solution.

- Specimen preparation
[Paraffin-embedded tissue sections]
Specimens may undergo histological disintegration or antigenic denaturation when subjected to highly concentrated fixative or prolonged fixing. Thus, in order to obtain an optimal fixing, maintaining tissue morphology and antigen activity, tissues which are as fresh as possible and small in size (about 1cm x 1cm x 0.5cm) should be used. The fixatives as shown below are recommended.

Fixing reagent	Fixing time
10% formalin or buffered formalin	24-48 hours
20% formalin	12-24 hours

- Section preparation
[Paraffin embedded tissue sections]
The cut sections should be 3-6 μm and placed on slides. When further treatments are to be done such as Antigen Recovery, Heat-Induced Epitope Retrieval (HIER) or trypsin treatment, the glass slides should be coated with an adhesive like 0.02% poly-L-lysine or silane for tissue sections.

[Control slides]
A positive control slide, negative control slide and reagent control slide are needed and processed in the same way as the unknown specimen slide to interpret staining results.

- Deparaffinization and Rehydration
- Treatment with xylene
 - Immerse the slides in xylene. After 3 minutes, take out and shake off the excessive xylene in the slides.
 - Repeat 1.(1) twice using fresh xylene.
 - Treatment with ethanol
 - Immerse slides in 100% ethanol. After 3 minutes, take out and shake off the excessive 100% ethanol in the slides.
 - Repeat 2.(1) once with fresh 100% ethanol.
 - Then, treat them twice with 95% ethanol in the same way as described above.
 - Washing
After excessive ethanol is shaken off, immerse slides in PBS for 5 minutes.
- Recommended Staining Procedures
- Quenching of endogenous peroxidase
 - Wipe areas around the sections on the slides carefully to remove excess solution.
 - Immerse them in 3% solution of hydrogen peroxide in absolute methanol for 10-15 minutes at room temperature.
 - Rinse them in fresh PBS for 3 times, each of 5 minutes duration.
 - Addition and reaction of Blocking reagent A(vial 1)
 - Wipe areas around the sections on the slides carefully.
 - Apply 2 drops of Blocking reagent A(vial 1) to specimen slide, positive control slide, negative control slide and reagent control slide respectively so as to provide a complete cover of the sections.
 - Incubate at room temperature for 60 minutes.
 - Rinse them in fresh PBS for 3 times, each of 5 minutes duration.
 - Addition and reaction of a mouse primary antibody
 - Wipe areas around the sections on the slides carefully.
 - Apply 2 drops of a mouse primary antibody to specimen slide, positive control slide and negative control slide respectively so as to provide a complete cover of the sections.
 - To the reagent control slide, apply two drops of negative control reagent (normal serum) in place of a mouse primary antibody.
 - Incubate them at room temperature or 4°C. (Follow the instructions for incubation time data designated in the package insert of primary antibody)
 - Rinse them in fresh PBS for 3 times, each of 5 minutes duration.
 - Addition and reaction of Blocking reagent B(vial 2)
 - Wipe areas around the sections on the slides carefully.
 - Apply 2 drops of Blocking reagent B(vial 2) to specimen slide, positive control slide, negative control slide and reagent control slide respectively so as to provide a complete cover of the sections.
 - Incubate at room temperature for 10 minutes.
 - Rinse them in fresh PBS for 3 times, each of 5 minutes duration.
 - Addition and reaction of Simple Stain Mouse MAX PO (M)(vial 3) (Universal Immuno-peroxidase Polymer).
 - Wipe areas around the sections on the slides carefully.
 - Apply 2 drops of Simple Stain Mouse MAX PO (M)(vial 3) to specimen slide, positive control slide, negative control slide and reagent control slide respectively so as to provide a complete cover of the sections.
 - Incubate at room temperature for 10 minutes.
 - Rinse them in fresh PBS for 3 times, each of 5 minutes duration.
 - Addition and reaction of chromogen/substrate reagent
 - Wipe areas around the sections on the slides carefully.
 - Apply 2 drops of the chromogen/substrate reagent to each slide so as to provide a complete cover of the sections. Incubate at room temperature for 5-20 minutes.
 - Rinse them in distilled water for 3 times, each of 5 minutes duration.
 - Counter-staining
 - Immerse them in the counterstain solution.
 - Wash them well with tap water.

8. Mounting
In case of alcohol soluble substrates like AEC, the tissue sections are mounted with water-based mounting media without further treatment. In case of alcohol insoluble substrates like DAB, they are mounted with permanent mounting media after washing with water, dehydrated in graded series of alcohol and cleared in xylene.

■ Interpretation of results

1. Microscopic observation
The slides are examined under a light microscope for a positive reaction. It is necessary to make comparison with three types of the control slides for interpreting staining results.
- Positive control slide
A specimen containing the target antigen which is processed in the same way as the unknown specimen.
- Negative control slide
A specimen not containing the target antigen which is processed in the same way as the unknown specimen

• Reagent control slide

The control specimen is used and processed in the same way as the test specimen except that negative control reagent is used instead of a primary antibody. If the slide is stained, it is probably due to non-specific reaction by non-specific protein binding.

7.STORAGE

Store in a dark place at 2-8°C.

8.LIMITATION

Tissue staining is dependent on the handling and processing of the tissue prior to staining. Improper fixation, freezing, thawing, washing, drying, heating, or sectioning may produce artifacts or false-negative results.

Results will not be optimal if old or unbuffered fixatives are used, or excessively heated during embedding or during attachment of sections to slides.

9.CONDITION FOR USE

M-Histofine® MOUSESTAIN KIT is designed for research use only and is not intended for therapeutic or diagnostic purposes. NICHIREI BIOSCIENCES INC.,

NICHIREI BIOSCIENCES sales agents and distributors will take no responsibility for **M-Histofine**® MOUSESTAIN KIT when used in a way which directly or indirectly violates local regulations or patents. Neither NICHIREI BIOSCIENCES nor its sales agents can be held responsible for any patent infringement which may occur as a result of improper use of the product.



Distributed in North America by:

COSMO BIO USA

2792 Loker Ave W, Suite 101
Carlsbad, CA 92010

TEL: 760-431-4600
FAX: 760-431-4604
email : info@cosmobiousa.com
web : www.cosmobiousa.com

10. TROUBLE SHOOTING

Problem	Possible cause	Solution
○ No staining or only weak staining results on the positive control slide and the unknown specimen slide.	<ol style="list-style-type: none"> 1. Drying-out of specimens during staining prior to addition of the reagents. 2. The embedding agent is not suitable, or paraffin is not thoroughly removed from paraffin-embedded sections. 3. Any trace amount of sodium azide present in the buffer inactivates the peroxidase, making the staining impossible. 4. Inadequate incubation of the enzyme and antibody. 	<ol style="list-style-type: none"> 1. Never allow the tissue to dry out. 2. Select a suitable embedding agent or remove paraffin thoroughly from sections embedded. 2. Change xylene or ethanol as the case may be. 3. Use sodium azide free buffer solution. 3. Change buffer solution. 4. Change stale chromogen/substrate reagent. 4. Blot off excess solution thoroughly at each stage. 4. Provide sufficient time for reaction with antibody. In particular, primary antibody should be incubated for the time period specified in the insert.
○ The unknown specimen slide is not stained while the positive control slide is stained.	<ol style="list-style-type: none"> 1. Antigen is denatured or masked during fixing or embedding process. 2. Antigen is decomposed by autolysis. 	<ol style="list-style-type: none"> 1. Some antigens are sensitive to fixation or embedding. So use less potent fixative and decrease the fixing time. 1. The pretreatment is required for some tissues, in order to reveal the antigen, such as Antigen Recovery, Heat-Induced Epitope Retrieval (HIER) or trypsin treatment. 2. Use tissues obtained by biopsy or surgery, whenever possible.
○ The backgrounds are intensively stained in all the slides.	<ol style="list-style-type: none"> 1. Endogenous peroxidase activity was not completely blocked. 2. Non -specific binding compositions are found. 3. Autolysis results in excessive antigens isolated in histological solutions. 4. Insufficient removal of paraffin. 5. Insufficient washing of antibody. 6. A high room temperature accelerates enzyme reactions. 7. Drying-out of specimens during staining after of the reagents. 	<ol style="list-style-type: none"> 1. Ensure that the procedure with 3% of hydrogen peroxidase added methanol. 2. Before adding primary antibody, treat with 10% normal goat serum. 3. Obtain fresh tissues whenever available. 4. Change xylene or ethanol as the case may be. 5. Ensure thorough washing of antibody. 6. Keep room temperature at 15 to 25°C 6. Shorten reaction time. 7. Never allow the tissue to dry out.
○ During the reaction, tissue sections come off from the slides.	<ol style="list-style-type: none"> 1. Some antigens require heat induced antigen retrieval procedure or prolonged reaction time with primary antibody, which may render the sections easily come off. 	<ol style="list-style-type: none"> 1. Mount tissue sections on slides coated with an adhesive such as 0.02% poly-L-lysine or silane.