



SYBR® Green Realtime PCR Master Mix -Plus-

QPK-212T 1mLx1
QPK-212 1mLx5

Store at -20°C, protected from light

Contents

[1]	Introduction
[2]	Components
[3]	Primer design
[4]	Detection
[5]	Specimens
[6]	Protocol
	1. Intercalator assay protocol using ABI PRISM® 7700
	2. Intercalator assay protocol using Roche LightCycler™
[7]	Troubleshooting
[8]	Related products

CAUTION

All reagents in this kit are intended for research purposes. Do not use for diagnosis or clinical purposes. Please observe general laboratory precaution and utilize safety while using this kit.

-LightCycler™ is a trademark of Idaho Technology, Inc. and Roche Molecular Systems, Inc.

-SYBR® is a registered trademark of Molecular Probes, Inc.

-ABI PRISM® is a registered trademark of Perkin-Elmer Corporation.

JAPAN

TOYOBO CO., LTD.
Tel(81)-6-6348-3888
www.toyobo.co.jp/e/bio
tech_osaka@toyobo.jp

CHINA

TOYOBO Bio-Technology, CO., LTD.
Tel(86)-21-58794900.4140

FOR RESEARCH USE ONLY. NOT FOR HUMAN OR DIAGNOSTIC USE.

[1] Introduction

Description

This product is an improved version of a Taq DNA polymerase-based 2 x master mix for real-time PCR, which contains all components, except for the primer. This reagent is applicable for highly specific intercalation assay with SYBR Green I. This reagent is effective for reducing non-specific amplifications or primer dimers.

Features

- Effective for reducing non-specific amplifications on SYBR[®] Green assay.
- Hot Start technology with anti-Taq DNA polymerase antibodies and additives enable high specificity and reproducible amplification.
- This reagent can be used in glass capillary systems (e.g., LightCycler, Roche Molecular Systems, Inc.).
- This reagent can be used in a passive reference system (e.g., ABI PRISM[®] 7700, Applied Biosystems, Inc.). The passive reference dye does not affect any other systems.

Notes:

This product cannot be used as a one-step RT-PCR reagent with reverse transcriptase. SYBR[®] Green Realtime PCR Master Mix (Code No. QPK-201) should be used instead.

[2] Components

This reagent includes the following components for 40 reactions (QPK-212T) and 200 reactions (QPK-212), 50 µl total reaction volume:

<QPK-212T>

SYBR [®] Green Realtime PCR Master Mix -Plus- Plus solution	1 ml 200 µl
---	----------------

<QPK-212>

SYBR [®] Green Realtime PCR Master Mix -Plus- Plus solution	1 ml x 5 1 ml
---	------------------

Notes:

This reagent can be stored at 4°C for 2 months and protected from light. For longer storage, this reagent should be kept at -20°C and protected from light.

[3] Primer design

Primers should be designed according to the following guidelines:

- Primer length: 20-30 mer
- GC content of primer: 40-60%
- Target length: ≤ 200 bp (optimally, ≤ 150bp)

Notes:

- Longer targets (>200 bp) reduce efficiency and specificity of amplification. The ideal target length range is 50-150 bp.
- Since detection in an intercalation assay is affected by non-specific amplification, primers should be as specific as possible.

[4] Detection

This reagent can be used in general detection devices, such as:

-LineGene (Bioer Technology co., ltd.)

This reagent can also be used in detection equipment using glass capillaries or passive reference, such as:

LightCycler (Roche Molecular Systems)

ABI PRISM® 7000, 7700, and 7900 (Applied Biosystems)

Note: The passive reference mode of detectors should be set at “ROX”.




Fig. 1 Melting curves by SYBR® Green Realtime PCR Master Mix -Plus- (left) and by a conventional kit (right).

[5] Specimen

1. cDNA

-Reverse transcription reactions from total or poly (A)⁺ RNA may be used directly, or after dilution, for real-time PCR. Purified cDNA by phenol/chloroform extraction and ethanol precipitation may also be used. Oligo dT and random primers are suitable for the reverse transcription reaction.

-Up to 20% of the synthesized cDNA solution from the ReverTra Ace® qPCR RT Kit (Code NO. FSQ-101) may be added to the PCR reaction solution directly, without purification.

2. Genomic DNA

-Purified DNA, which would be used for general PCR, is also suitable for real-time PCR. In the case of mammalian genomic DNA, 1-10 ng genomic DNA is sufficient for real-time PCR.

[6] Protocol

1. Intercalation assay protocol using ABI PRISM® 7700

The following is an intercalator assay protocol to be used with ABI PRISM® 7700. For other detection devices, this protocol may require modification depending on each instruction manual.

(1) Preparation of reaction solution

Component	Volume	Final Concentration
PCR grade water	11 µl	
SYBR® Green Realtime PCR Master Mix	25 µl	1x
Plus solution	5 µl	
10pmol/µl (10 µM) Primer #1	2 µl	0.4 µM
10pmol µl (10 µM) Primer #2	2 µl	0.4 µM
Template DNA	5 µl	
Total volume	50 µl	

Notes

-The primer concentration can be further optimized, if needed. The optimal range for the primers is 0.2-0.6 µM. In the case of commercially available primers, recommended conditions from those companies should be used.

(2) Cycling conditions

The following condition is recommended. Initially, a 3-step cycle should be used.

<3-step cycle>		
Pre-denaturation:	95°C, 1 min.	
Denaturation:	95°C, 15 sec.	40 cycles
Annealing:	55-65°C, 15 sec.	
Extension:	72°C, 45 sec. (data collection)	
Melting curve analysis		
<2-step cycle>		
Pre-denaturation:	95°C, 1 min.	
Denaturation:	95°C, 15 sec.	40 cycles
Extension:	60°C, 1 min. (data collection)	
Melting curve analysis		

Notes

- The annealing temperature in 3-step cycle should be set to 55-65°C, depending of the primer T_m value.
- The pre-denaturation condition described above is sufficient for inactivation of the anti-Taq DNA polymerase antibodies used in Hot Start PCR. To prevent unexpected and inappropriate results, do not prolong the pre-denaturation period. Fifteen seconds is also sufficient for denaturation during each cycle.
- Data collection step should be longer than 30 sec.

2. Intercalation assay protocol using Roche LightCycler™

The following is an intercalator assay protocol to be used with the Roche LightCycler™. In the case of other detection devices, this protocol should be modified accordingly.

(1) Preparation of reaction solution

Component	Volume	Final Concentration
PCR grade water	3.6 µl	
SYBR® Green Realtime PCR Master Mix	10 µl	1x
Plus solution	2 µl	
10pmol/µl (10 µM) Primer #1	1.2 µl	0.6 µM
10pmol/µl (10 µM) Primer #2	1.2 µl	0.6 µM
Template DNA	2 µl	
Total volume	20 µl	

Notes

-The primer concentration can be further optimized, if needed. The optimal range for primers is 0.4-0.8 µM. In the case of commercially available primers, recommended conditions from each manual should be followed.

(2) Cycling condition

The following condition is recommended:

<3-step cycle>		
Pre-denaturation:	95°C, 30 sec.	} 40 cycles
Denaturation:	95°C, 5 sec.	
Annealing:	55-65°C, 10 sec.	
Extension:	72°C, 15 sec. (data collection)	
Melting curve analysis		

Notes

- The annealing temperature can be set to 55-65°C, depending on the primer T_m.
- The annealing time should be set for 5-20 seconds. Longer annealing time results in increased efficiency, and a shorter time decreases non-specific amplification.
- The pre-denaturation condition described above is sufficient for inactivation of the anti-Taq DNA polymerase antibodies used in Hot Start PCR. To prevent unexpected and inappropriate results, do not prolong the pre-denaturation period. Five seconds is also sufficient for denaturation during each cycle.
- Data collection step should be longer than 10 sec.
- If commercially available primers or probes are employed, the recommended conditions from each company should be used.

[7] Troubleshooting

Symptom	Cause	Solution
No amplification	Incorrect mode setting of detector for the fluorescent dye.	Confirm the detector setting.
	Incorrect setting for data collection.	Confirm the data collection setting.
	Incorrect setting for sample position.	Reposition the sample tubes.
	Inappropriate concentration of primers.	Optimize the concentration of primers according to the instructions (see [6]).
	Inappropriate design of primers.	Confirm the specificity and T _m values of the primers.
	Inappropriate cycle conditions.	Confirm the primer T _m value. A lower annealing temperature increases amplification efficiency. For GC rich targets, an elongated denaturation time might be effective.
	Low purity or quality of samples	Check the purity or quality of sample DNA.
Variation in detection	Failure or malfunction of the device	Check the device.
	Low quality sample DNA.	Repurify the sample DNA by phenol/chloroform extraction and ethanol precipitation, or other method.
	Inappropriate concentration of primers.	Optimize the concentration of primers according to instructions (see [6]). A lower annealing temperature reduces variations. For GC rich targets, elongated denaturation time might be effective.
	Inappropriate design of primers.	Confirm specificity and T _m values of the primers.
	Inappropriate cycle conditions.	Confirm the T _m values of the primers.
	Variation of dispensed volume	Increase the reaction volume
Signals in blank reactions	Contamination of amplicons or sample DNAs.	Use fresh PCR grade water, and remake the primer solution and master mix.
	Detection of non-specific amplification.	Optimize the primer and cycle conditions.

[8] Related products

Product name	Package	Code No.
High efficient cDNA synthesis kit for real-time PCR ReverTra Ace[®] qPCR RT Kit	200 reactions	FSQ-101
High efficient reverse transcriptase ReverTra Ace[®]	10,000 U	TRT-101
RNase inhibitor	2,500 U	SIN-201
Real-time PCR master mix for probe assay Realtime PCR Master Mix	1 mL x 1 1 mL x 5	QPK-101T QPK-101
Real-time PCR master mix for SYBR [®] Green assay SYBR[®] Green Realtime PCR Master Mix	1 mL x 1 1 mL x 5	QPK-201T QPK-201
One-step real-time PCR master mix for probe assay RNA-direct[™] Realtime PCR Master Mix	0.5 mL x 2 0.5 mL x 5	QRT-101T QRT-101
One-step real-time PCR master mix for SYBR [®] Green assay RNA-direct[™] SYBR[®] Realtime PCR Master Mix	0.5 mL x 2 0.5 mL x 5	QRT-201T QRT-201



NOTICE TO PURCHASER: LIMITED LICENSE

Use of this product is covered by one or more of the following US patents and corresponding patent claims outside the US: 5,079,352, 5,789,224, 5,618,711, 6,127,155, 5,677,152, 5,773,258, 5,407,800, 5,322,770, 5,310,652, 5,994,056, 6,171,785, and claims outside the US corresponding to US Patent No. 4,889,818. The purchase of this product includes a limited, non-transferable immunity from suit under the foregoing patent claims for using only this amount of product for the purchaser's own internal research. No right under any other patent claim (such as apparatus or system claims in US Patent No. 6,814,934) and no right to perform commercial services of any kind, including without limitation reporting the results of purchaser's activities for a fee or other commercial consideration, is conveyed expressly, by implication, or by estoppel. This product is for research use only. Diagnostic uses under Roche patents require a separate license from Roche. Further information on purchasing licenses may be obtained by contacting the Director of Licensing, Applied Biosystems, 850 Lincoln Centre Drive, Foster City, California 94404, USA.

JAPAN

TOYOBO CO., LTD.
Tel(81)-6-6348-3888
www.toyobo.co.jp/e/bio
tech_osaka@toyobo.jp

CHINA

TOYOBO Bio-Technology, CO., LTD.
Tel(86)-21-58794900.4140

FOR RESEARCH USE ONLY. NOT FOR HUMAN OR DIAGNOSTIC USE.